

MGT-N/P 4 GAS DETECTOR DOCKING STATION

Full management of up to 4 Senko MGT-N/P detectors at the same time



UP TO 4 DEVICES AT ONCE

MANAGEMENT OF

- Calibration
- Bump Test
- Event Logging (USB)
- Setpoint adjustment

PC SOFTWARE

LED INDICATION OF STATUS

AUTOMATIC GAS VENTING

OPTIMIZATION OF GAS CONSUMPTION

RECHARGEABLE BATTERY OR POWER

Configuration

Calibration interval: 0~365 days

Bump interval: 0~365 days

Low / High Alarm Setting
 - H₂S : 0 ~ 100.0
 - CO : 0 ~ 300
 - O₂ : 0 ~ 30.0
 - LEL : 0 ~ 100.0

STEL / TWA Alarm Setting
 - H₂S : 0 ~ 100.0
 - CO : 0 ~ 300

Detector User ID

TWA Method
 - ACGIH : Accumulation Average
 - OSHA : Movement Average

TWA
 - Default : 8 hours average

STEL
 - Default : 15 minutes average

Safe Display : In the safe area, SAFE will be displaced.
 Dock Lock : Lock the bump / calibration
 Maintenance Notification
 Latching Alarms : Until the users push the button, the alarm will be activated continuously.
 Auto Zero : When the device is turned on, the zero calibration will be activated.
 Lel by volume : 100%LEL Standard
 Off Lock : Prevent users from turning off the device.

Choose the MGT firmware

Initiate the firmware upload.

**RECHARGEABLE LI-ION
 BATTERY FOR UP TO 1,000
 BUMP TESTS**

